

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D2046627	<b>(X3) Date Survey Completed</b>  03/24/2025
<b>Name of Provider or Supplier</b>  Advanced Urgent Care Center	<b>Street Address, City, State</b>  1980 N Roosevelt Blvd, Key West, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced CLIA recertification survey was conducted at Advanced Urgent Care Center from March 13, 2025 to March 24, 2025. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiency:
<b>D2015</b>	<p><b>TESTING OF PROFICIENCY TESTING SAMPLES</b> CFR(s): 493.801(b)(5)(6)</p> <p>(b)(7) PT is required for only the test system, assay, or examination used as the primary method for patient testing during the PT event.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to retain copies of the instrument test results for Chemistry proficiency event #3 for the year 2023, one out of six events reviewed 2023 (events #1, #2 and #3) and 2024 (events #1, #2 and #3). Findings included: 1-Review of the American Proficiency Institute (API) test records showed that on 9/11/2023 the laboratory did not maintain a copy of the instruments test results from the Triage-Alere analyzer for the 3rd event in 2023. 2- Review of the laboratory Proficiency Testing procedure manual revealed in step "11. All instrument tapes, worksheets, attestation and the reviewed survey results will be kept readily available for two years." Additional instructions stated in bold print "NOTE: The following data must be retained for two years: ...The instrument tapes /printout..." 3- Interview on 03/13/2025 at approximately 12:05 PM, the laboratory Technical Consultant confirmed that the laboratory did not keep the records of Triage-Alere analyzer for this event.</p>