

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D2048999	<b>(X3) Date Survey Completed</b>  08/20/2018
<b>Name of Provider or Supplier</b>  Pediatric Associates	<b>Street Address, City, State</b>  9655 Nw 41st St, Doral, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2121</b>	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to receive a passing proficiency test (PT) score for the third testing event of 2017 for the specialty of Hematology. Findings: Review of PT score from American Proficiency Institute for the third event of 2017 showed a score of 73% for White Blood Cells (WBC) Differential. During an interview on 8/20/18 at 1:50 PM, the Laboratory Manager confirmed the laboratory had failed proficiency testing for WBC differential.</p>