

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D2056098	<b>(X3) Date Survey Completed</b>  04/10/2023
<b>Name of Provider or Supplier</b>  Martin Dermatology Pl	<b>Street Address, City, State</b>  2645 Sw 37th Ave Ste 505, Miami, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A recertification survey conducted on 04/10/2023 found the MARTIN DERMATOLOGY PL clinical laboratory not in compliance with 42 CFR Part 493, Requirements for Laboratories.
<b>D3011</b>	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on observation, record review and interview, the laboratory failed to store the following flammable reagents according to manufacturer instructions: 100% Reagent Alcohol, Xylene Substitute XS-3 and Eosin-Y alcoholic 0.25%. Findings included: During a tour of the laboratory on 04/10/2023 at 9:45 AM, the surveyor found that the laboratory stored in an unlocked cabinet that it is not fireproof the following reagents: one gallon of 100% reagent alcohol, one gallon of XS-3 and one gallon of Eosin-Y alcoholic 0.25%. Review of the STATLAB Safety Data Sheet for those products revealed the following: -100% Reagent Alcohol to "Store locked up". - XS-3 and Eosin-Y alcoholic 0.25% to "Keep in a fireproof place". During an interview on 04/10 /2023 at 9:50 AM, the risk manager confirmed that the laboratory failed to store the reagents listed above as per manufacturer instructions.</p>