

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2059340	(X3) Date Survey Completed 01/04/2019
Name of Provider or Supplier Quest Diagnostics Clinical Laboratories Inc	Street Address, City, State 9220 Estero Park Commons Blvd Ste 1, Estero, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2005	<p>ENROLLMENT CFR(s): 493.801(a)(4)</p> <p>Authorize the proficiency testing program to release to HHS all data required to-- (i) Determine the laboratory's compliance with this subpart; and (ii) Make PT results available to the public as required in section 353(f)(3)(F) of the Public Health Service Act.</p> <p>This STANDARD is not met as evidenced by: Based on observation, lack of documentation, and proficiency testing result information from the laboratory unit, the laboratory failed to authorize the proficiency testing program to release information. Findings include: The surveyor attempted to look up proficiency testing results for the laboratory before the survey, and there were no results available on the Center for Medicare and Medicaid Services (CMS) reports. The laboratory unit verified that they did not have this laboratory on their College of American Pathologists (CAP) proficiency testing list. Review of proficiency testing results on 01/04/19 at 10:00 a.m. revealed that the laboratory was enrolled in proficiency testing with the CAP with successful scores over the past two years. During an interview with the laboratory supervisor at 11:30 a.m. on 01/04/19, she was not aware that results were not authorized to be released.</p>