

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2062745	(X3) Date Survey Completed 02/05/2026
Name of Provider or Supplier Nosky Pa	Street Address, City, State 512 Cypress Pkwy, Kissimmee, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5400	<p>ANALYTIC SYSTEMS CFR(s): 493.1250</p> <p>Each laboratory that performs nonwaived testing must meet the applicable analytic systems requirements in 493.1251 through 493.1283, unless HHS approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub.7), that provides equivalent quality testing. The laboratory must monitor and evaluate the overall quality of the analytic systems and correct identified problems as specified in 493.1289 for each specialty and subspecialty of testing performed.</p> <p>This CONDITION is not met as evidenced by: Based on observation, record review, review of the procedure manual, and interview, the laboratory failed to use non-expired reagents in their Hematoxylin (H) and Eosin stain for staining patient's histopathology slides on 09/18/2025 (See D5417), and the laboratory failed to document the acceptability of the Hematoxylin and Eosin (H&E) control slide for two days (11/20/2025, 12/18/2025) of the quality control (QC) log reviewed from 11/13/2025 to 02/05/2025 (See D5601). These are uncorrected deficiencies from the survey on 10/08/2025.</p>
D6108	<p>LABORATORY TECHNICAL SUPERVISOR CFR(s): 493.1447</p> <p>The laboratory must have a technical supervisor who meets the qualification requirements of 493.1449 of this subpart and provides technical supervision in accordance with 493.1451 of this subpart.</p> <p>This CONDITION is not met as evidenced by:</p>

Based on review of the procedure manual, record review, and interview, the Technical Supervisor (A) failed to identify problems in the documentation of quality control and maintenance records from 11/13/2025 to 02/05/2025. (See D6123)