

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2063340	(X3) Date Survey Completed 10/08/2024
Name of Provider or Supplier Gulf Coast Pain Consultants Plc	Street Address, City, State 4901 Marketplace Rd, Pensacola, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Gulf Coast Pain Consultants PLLC on 10/08/2024. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D3011	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview it was determined safety procedures to ensure protection from chemical hazards were not followed for four of four unopened bottles of Acetone. Findings include: During a tour of the lab on 10/08/24 at 10:55 AM, four unopened bottles of Acetone were observed on the open shelf under the laboratory hood. The safety label for the Acetone bottles indicated the contents were flammable and should be stored appropriately. The General Supervisor confirmed on 10/08/24 at 10:55 AM Acetone should be stored in the laboratory Safety Cabinet. The Technical Consultant confirmed via phone interview on 10/08/24 at 11:30 AM Acetone was to be stored in the laboratory Safety Cabinet.</p>