

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2075325	(X3) Date Survey Completed 01/24/2019
Name of Provider or Supplier Skin And Cancer Associates, Llp	Street Address, City, State 21020 State Rd #120, Boca Raton, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with testing personnel (TP) A, the laboratory failed to document the annual competency assessment on 1 out of 6 (A, B, C, D, E, F) testing personnel for 2017 Findings include: Review of employee documentation showed that the laboratory failed to have documentation of the annual competence evaluation for the year 2017 for TP E During an interview on 01/24/2019 at 11:00 AM, with the TP A, he confirmed that there was no annual competency assessment documented for TP E for the period of reference.</p>