

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D2080975	<b>(X3) Date Survey Completed</b>  07/10/2025
<b>Name of Provider or Supplier</b>  Florida Physician Specialists Llc	<b>Street Address, City, State</b>  710 Lomax St, Jacksonville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	At the time of the announced, onsite recertification survey, Florida Physician Specialists was found to not be in compliance with the CLIA laboratory requirements of 42 CFR 493.
<b>D2009</b>	<p><b>TESTING OF PROFICIENCY TESTING SAMPLES</b> CFR(s): 493.801(b)(1)</p> <p>(b)(1) The individual testing or examining the samples and the laboratory director must attest to the routine integration of the samples into the patient workload using the laboratory's routine methods.</p> <p>This STANDARD is not met as evidenced by: Based on review of proficiency testing records and interview with staff, the Laboratory Director or designee failed to sign proficiency test attestation statements for 9 of 9 testing events reviewed. The findings include: Review of American Proficiency Institute (API) records showed the Laboratory Director or designee failed to sign API's attestation statements attesting to the routine integration of the proficiency samples into the patient workload using the laboratory's routine methods for the following testing events: a. 2024 - Chemistry Event 1, Event 2, and Event 3 b. 2024 - Microbiology Event 1, Event 2, and Event 3 c. 2025 - Chemistry Event 1, Event 2 d. 2025 - Microbiology Event 1 During interview on 7/10/2025 at 11:40 AM, testing person #1 confirmed the attestation statements listed above were not signed by the Laboratory Director or designee. .</p>