Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2089479	(X3) Date Survey Completed 09/16/2020
Name of Provider or Supplier Uf Health North	**Street Address, City, State** 15255 Max Leggett Pkwy, Jacksonville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	At the time of the announced, on-site recertification survey, UF Health North, was found to be NOT in compliance with the CLIA laboratory requirements of 42 CFR 493. .
D2094	ROUTINE CHEMISTRY CFR(s): 493.841(e) (1) For any unsatisfactory analyte or test performance or testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure. (2) For any unacceptable analyte or testing event score, remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event. This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, there was no documentation to indicate that the laboratory took corrective action for the unsatisfactory score in proficiency testing received in 2019. Findings include: The laboratory received an unsatisfactory score of 67% for Parathyroid Hormone in the American Proficiency Institute 1st Chemistry event of 2019 . There was no documentation to indicate what had happened, if the laboratory had investigated, or that corrective action was taken. During an interview with the Laboratory Supervisor. at 11:00am on 9/16/20, it was said that corrective action had been performed but the documentation had been misplaced.