

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2089763	(X3) Date Survey Completed 07/31/2018
Name of Provider or Supplier Uf Health Dermatology Deerwood Park	Street Address, City, State 10475 Centurion Pkwy N Ste 106, Jacksonville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with laboratory staff, the facility failed to have potassium hydroxide that was not expired for 3 months in 2018. The findings include: During observation on 07/31/18 at 8:50 AM, the potassium hydroxide located in the patient exam rooms had an expiration date of April 2018. The interview on 7/31/18 at 8:50am with the laboratory testing personnel confirmed the potassium hydroxide was expired.</p>