Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 10D2089763	**(X3) Date Survey Completed** 09/27/2022
Name of Provider or Supplier Uf Health Dermatology Deerwood Park	**Street Address, City, State** 10475 Centurion Pkwy N Ste 106, Jacksonville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	At the time of the announced, onsite recertification survey, UF Health Dermatology Deerwood Park was found to NOT be in compliance with the CLIA laboratory requirements of 42 CFR 493.
D5413	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (1) Water quality. (2) Temperature. (3) Humidity. (4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports. This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to document room temperature and humidity in the laboratory for 2 (2020-2022) out of 2 years reviewed. The findings include: The laboratory was unable to provide documentation showing room temperature and humidity of the laboratory was being monitored each day of patient testing. During an interview on 9/27/22 at 11:27 AM the Laboratory Director confirmed that room temperature and humidity were not being recorded.