

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2097728	(X3) Date Survey Completed 07/16/2019
Name of Provider or Supplier Genesiscare Usa Of Florida Llc DbA	Street Address, City, State 20601 E Dixie Hwy Ste 330, Aventura, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at 21st Century Oncology DbA Integrated Oncology of South Florida on 7/16/2019. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D2121	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of Medical Laboratory Evaluation (MLE) proficiency testing records and interview with Office Consultant (OC), the laboratory failed to score at least 80 % on Red Blood Count (RBC) for 1 (2nd event of 2017) out of 7 events for Hematology reviewed. Findings include: Review of MLE proficiency records revealed a score of 60% for RBC in the 2nd event of 2018. During an interview on 07 /16/2019 at 10:30 AM, the OC confirmed the failed proficiency testing score.</p>
D2127	<p>HEMATOLOGY CFR(s): 493.851(d)</p> <p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of Medical Laboratory Evaluation (MLE) proficiency testing (PT) results and interview with Office Consultant (OC) the laboratory failed to submit PT results in the specified timeframe for 2nd event of 2019 for the specialty of</p>

Hematology resulting in a score of 0% for all hematology analytes. Findings include: Review of MLE Hematology PT records revealed that the laboratory failed to submit the results for the second event of 2018 on time resulting in a score of 0 % scores for Hematology, Cell Identification or White Blood Differential, Red Blood Cell, Hematocrit, Hemoglobin, White Blood Cells and Platelets. During an interview on 7 /16/2019 at 10:30 AM, the OC confirmed that the laboratory failed to submit on time the results for the 2nd event of 2019.