

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2098736	(X3) Date Survey Completed 01/30/2019
Name of Provider or Supplier Saltzman Tanis Pittell Levin & Jacobson Inc D/B/A	Street Address, City, State 379 N Congress Ave, Boynton Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2020	<p>BACTERIOLOGY CFR(s): 493.823(a)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, the laboratory failed to receive a score of 80 percent for bacteriology for the first of three testing events of 2018. Findings include: Review of proficiency testing records for the past two years on 01/30/19 revealed that the laboratory had unacceptable scores for two of five proficiency testing samples and received an overall score of 60% for bacteriology for the first testing event of 2018. During an interview with the laboratory manager at 1:30 p.m. on 01/30/19, she confirmed that they were aware of the unacceptable score and had completed corrective action.</p>