

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2101438	(X3) Date Survey Completed 02/06/2025
Name of Provider or Supplier Park Avenue Dermatology North	Street Address, City, State 15255 Max Leggett Pkwy Ste 5100, Jacksonville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	At the time of the announced, onsite recertification survey, Park Avenue Dermatology North was found to NOT be in compliance with the CLIA laboratory requirements of 42 CFR 493.
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to perform competency evaluations on 2 of 2 Testing Personnel who perform potassium hydroxide (KOH) and Scabies testing in the subspecialties of Mycology and Parasitology for 2 of 2 (2023-2024) years reviewed. Findings include: The review of employee competency records showed that no competency evaluations were conducted for Testing Personnel #C and #D, who perform KOH (potassium hydroxide) and scabies testing, in 2023 and 2024. During the interview on 02/6/2025 at 10:08 AM with the Mohs technician, it was confirmed that the laboratory had not performed competency evaluations for Testing Person #C and #D. .</p>
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p>

This STANDARD is not met as evidenced by:

Based on observation and staff interview, the laboratory failed to ensure the PMS Fungal/Tzanck stain was not expired prior to patient testing. Findings include:

Observations made during a tour of the laboratory on 2/6/25 at 9:58 AM, showed one bottle of "PMS Fungal/Tzanck" stain with lot number K21551 expired 5/31/2023, and a second bottle of "PMS Fungal/Tzanck" with lot number K229C4 expired 9/30/24. During an interview on 2/6/25 at 10:10 AM with the Mohs tech, it was confirmed the two bottle of stain were expired.