

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2103383	(X3) Date Survey Completed 11/07/2023
Name of Provider or Supplier John W Mcdonald Md Faad Pa	Street Address, City, State 1770 37th St, Vero Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	Recertification survey was conducted on November 7, 2023. John W McDonald MD FAAD PA clinical laboratory was not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on interview, and review of peer review records, the laboratory failed to maintain documentation to verify accuracy of the reading and interpretation of the Hematoxylin and Eosin (H&E) stain at least twice annually for 2022. The findings include: Review of peer review records revealed that each year's peer review was stored in separate envelopes. There was no envelope for 2022. On 11/07/2023 at 10:50 AM, the Laboratory Director stated he did not have any peer review for 2022.</p>