

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2103965	(X3) Date Survey Completed 07/28/2023
Name of Provider or Supplier Orchid Dermatology Llc	Street Address, City, State 5301 4th Avenue Cir E, Bradenton, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was conducted on July 28, 2023. Orchid Dermatology LLC clinical laboratory was not in compliance with 42 CFR 493, Requirements for Laboratories. The following standard level deficiency was cited:
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of peer review records and interview, the laboratory failed to have documentation to verify accuracy of the reading and interpretation of the Hematoxylin and Eosin (H&E) stain at least twice annually for one (2022) of two years reviewed (2021-2023). Findings included: Review of the Mohs Quality Control Evaluation forms showed peer review was done on 04/13/2021, 11/02/2021, 09/13/2022, and 01/12/2023. On 07/28/2023 at 11:40 AM, the Mohs Technician confirmed she could not locate documentation of the second peer review done in 2022.</p>