Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2103965	(X3) Date Survey Completed 07/23/2025
Name of Provider or Supplier Orchid Dermatology Llc	**Street Address, City, State** 5301 4th Avenue Cir E, Bradenton, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Orchid Dermatology LLC on 7/23/25. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D3011	FACILITIES CFR(s): 493.1101(d) Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials. This STANDARD is not met as evidenced by: Based on observation and interview, the laboratory failed to safely store two of two Hazardous Waste containers to ensure protection from chemical hazards. Findings included: 1. During a tour of the laboratory on 7/23/25 at 10:30 a.m., two plastic containers labeled as Hazardous Waste with a Flammable Liquid caution sticker on them were observed on top of the laboratory safety flammable cabinet. 2. The Histology tech confirmed on 7/23/25 at 10:50 a.m., the two Hazardous Waste containers should be stored in the laboratory safety flammable cabinet, but there was not enough room.