

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2104407	(X3) Date Survey Completed 06/23/2020
Name of Provider or Supplier Beach Care Pllc	Street Address, City, State 1000 5th Street Suite 200-Ys, Miami Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An onsite complaint visit, #2020003843 was conducted on 06/23/20 at Beach Care PLLC. The laboratory was not in compliance with 42 CFR 493, Requirements for Clinical Laboratories.
D2093	<p>ROUTINE CHEMISTRY CFR(s): 493.841(d)</p> <p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on CLIA records reviewe and interview, the laboratory failed to notify to the proficiency testing Agency that the laboratory was not testing before the due date to submit the proficiency testing results for 1 out of 1 event for Routine Chemistry (3rd event 2019). Findings include: Record review revealed the laboratory was enrolled in proficiency testing with American Academy of Family Physician. Review of CLIA Casper report 096 CLIA Application and Summary revealed a score of 0% for Arterial blood pH (PH Blood Gas), Partial pressure Oxygen (PO2), Partial pressure Carbon dioxide (PCO2), Chloride (CL), Creatine Kinase Total, Creatine Kinase (CK) Iso, Creatine, Glucose, Potassium (K), Serum Sodium(NA), Blood Urea Nitrogen (BUN) and an overall score of 0 % for Routine Chemistry for the 3rd event of 2019. During an interview on 6/23/2020 at 2:30 PM, the owner confirmed that the laboratory failed to notify to the proficiency testing agency that the laboratory was not testing within the timeframe for the event of reference.</p>
D2123	<p>HEMATOLOGY CFR(s): 493.851(c)</p> <p>Failure to participate in a testing event is unsatisfactory performance and results in a</p>

score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if-- (1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results; (2) The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and (3) The laboratory participated in the previous two proficiency testing events.

This STANDARD is not met as evidenced by:

Based on CLIA records review and interview, the laboratory failed to notify to the proficiency testing Agency that the laboratory was not testing before the due date to submit proficiency testing records for 1 out of 1 event for Hematology (3rd event 2019). Findings include: Record review revealed the laboratory was enrolled in proficiency testing with American Academy of Family Physician. Review of CLIA Casper report 096 CLIA Application and Summary revealed a score of 0% for platelet and an overall score of 0 % for the Hematology 3rd event of 2019. During an interview on 6/23/2020 at 2:30 PM, the owner confirmed that the laboratory failed to notify to the proficiency testing agency that the laboratory was not testing within the timeframe for the event of reference.