| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number<br><br>10D2105402 | (X3) Date Survey Completed<br><br>05/23/2019 |
|---|---|---|
| **Name of Provider or Supplier**<br><br>Aqua Dermatology Of Florida Pa | **Street Address, City, State**<br><br>9905 Old St Augustine Rd Bldg 200, Jacksonville, FL | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---|---|
| **D5417** | TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT<br>CFR(s): 493.1252(d)<br><br>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.<br><br>This STANDARD is not met as evidenced by:<br>Based on record review and staff interview, the laboratory used expired Eosin-Y reagent (used for staining patient tissue pathology slides) for 5 of 6 months reviewed in 2018. Findings included: Record review of the "H & E Permanent Stain Line" log for 2018 showed the reagent name "Eosin-Y" with lot number 162104 had an expiration date of 8/3/2018. MOHs testing of 17 patients was performed on 8/21/18, 9 /18/18, and 10/16/18. MOHs testing was performed on 18 patients on 11/20/18 and 12 /18/18. The documentation showed the expired Eosin-Y was used on all five dates. An interview was done on 5/23/19 at 8:30am with the Histotechnologist. They confirmed the Eosin-Y was expired when used to stain patient slides. |