

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2106298	(X3) Date Survey Completed 10/26/2021
Name of Provider or Supplier Tarantola Dermatology Inc	Street Address, City, State 525 Brent Lane, Pensacola, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A CLIA recertification survey was conducted on October 26, 2021. Tarantola Dermatology clinical laboratory was not in compliance with 42 CFR 493, Requirements for Clinical Laboratories.
D5407	<p>PROCEDURE MANUAL CFR(s): 493.1251(d)</p> <p>Procedures and changes in procedures must be approved, signed, and dated by the current laboratory director before use.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory staff, the laboratory director failed to sign the policy and procedure manual since 2019. Findings included: Review of the policy and procedure manual revealed that it was last signed May 7, 2019. During interview with laboratory staff at 11:45 on 10/26/2021, it was confirmed the policy and procedure manual had not been signed since 2019.</p>