

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2108487	(X3) Date Survey Completed 07/10/2019
Name of Provider or Supplier Aqua Dermatology Of Florida Pa	Street Address, City, State 121 S State Road 7, Wellington, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Water's Edge Dermatology on 07/10/2019. The laboratory was not in compliance with Code of Federal Regulations (CFR), Part 493, requirements for clinical laboratories.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, the laboratory did not verify the accuracy of mycology testing twice annually. Findings include: Review of verifications of potassium hydroxide (KOH) preparations on 07/10/2019 revealed that the laboratory verified the accuracy of KOH preps only once in 2018. During an interview with the laboratory histotechnologist at 9:45 a.m. on 07/10/2019, she confirmed that they must have skipped one verification.</p>