

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D2111405	<b>(X3) Date Survey Completed</b>  07/25/2018
<b>Name of Provider or Supplier</b>  Florida Cancer Specialists & Research Institute	<b>Street Address, City, State</b>  3730 7th Terr Ste 101b, Vero Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2122</b>	<p>HEMATOLOGY CFR(s): 493.851(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the facility failed to receive a passing proficiency test score for the third testing event of 2016 for the specialty of Hematology. Findings: Review of the American Proficiency Institute testing results for the third event of 2016 showed a score of 0% for White Blood Cell Differential, Granulocytes, Lymphocytes, and Monocytes. During an interview on 7/25/18 at 12:30 PM, the Laboratory Supervisor confirmed the laboratory had failed the proficiency testing due to clerical error.</p>