

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2116797	(X3) Date Survey Completed 06/25/2026
Name of Provider or Supplier Erica Smithberger Md Pl DbA	Street Address, City, State 2010 Thonotosassa Rd, Plant City, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Erica Smithberger MD PL DBA Clayton Dermatology Group on 06/25/2026. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Standard deficiencies cited are as follows:
D3011	FACILITIES CFR(s): 493.1101(d) Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials. This STANDARD is not met as evidenced by: Based on observation, interview, and review safety data sheets (SDS), the laboratory failed to ensure protection from chemical hazards used for Histopathology testing of Hematoxylin and Eosin (H&E) stain from 06/05/2024 to 06/25/2026. Findings: 1. During a tour of the laboratory on 06/25/2026 at 10:25 AM, no fume hood was observed to provide protection from respiratory chemical hazards. 2. The SDS sheets for Xylene Substitute, Reagent Alcohol 100%, and Reagent Alcohol 95%, all included hazard statements regarding inhalation precautions and to not breathe "dust/ fume/ gas/ mist/vapors/spray". 3. The Histology Tech confirmed on 06/25/2026 at 10:25 AM there was no hood or any other method to ensure respiratory protection from chemical hazards used for H & E staining.