

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2119256	(X3) Date Survey Completed 10/30/2018
Name of Provider or Supplier Bassan & Bloom Mds Pl	Street Address, City, State 1701 Ne 164th St Ste 304, North Miami Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with testing personnel (TP), the laboratory failed to document the initial training and six months competency assessment on 1 out of 1 testing personnel for 2017 and 2018 period. Findings include: Review of employee documentation showed that the laboratory failed to have documentation of the initial training on testing person for the year 2017 and no six months competency assessment on 2018. During an interview on 10/30/2018 at 10:30 AM, with the TP, she confirmed that there was no initial and no six months competency assessment documented for the period of reference.</p>