

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2119907	(X3) Date Survey Completed 06/11/2019
Name of Provider or Supplier Leon Medical Centers, Llc	Street Address, City, State 8881 Nw 18th Terrace Ste 102, Doral, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Leon Medical Centers Inc on 6/11/2019. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D2121	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of American Proficiency Institute (API) proficiency testing records and interview with Testing Person #1, the laboratory failed to score at least 80 % on White Blood cell Diferrential (WBC DIFF) for 1 (2nd event of 2018) out of 6 events for Hematology reviewed. Findings include: Review of API proficiency records revealed a score of 33 % for WBC DIFF in the 2nd event of 2018. During an interview on 06/11/2019 at 10:30 AM, the TP # 1 confirmed the failed proficiency testing score.</p>