

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2121836	(X3) Date Survey Completed 12/04/2018
Name of Provider or Supplier Clarity Genetics Llc	Street Address, City, State 5231 Nw 33rd Ave, Fort Lauderdale, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with general supervisor (GS), the laboratory failed to perform the six months competency assessment for 1 out of 1 testing personnel (TP) for 2017 year. Findings include: Review of employee competency policy revealed that the requirement for the 6th month evaluation in the first year of employment was missing. Review of employee documentation showed that the laboratory failed to have documentation of the 6th month evaluation on testing person for the year 2017. During an interview on 12/03/2018 at 10:30 AM, with the GS, he confirmed that the competency assessment policy missed the requirement for the 6th month assessment during 1st year of employment and that there was no six months competency assessment documented for the period of reference for TP.</p>