Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2124280	(X3) Date Survey Completed 02/26/2019
Name of Provider or Supplier Clinical Discovery Institute Llc	**Street Address, City, State** 7901 Riviera Blvd Suite 100, Miramar, FL	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2007	TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1) The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods This STANDARD is not met as evidenced by: Based on review of CMS 209 Laboratory Personnel Report, review of College of American Pathologist (CAP) proficiency testing (PT) records and interview with Testing Personnel (TP) A the laboratory failed to have all testing personnel (TP) rotate through the testing of PT for the hematology specialty for 2 out of 2 years (2017 - 2018) reviewed. Findings include: Review of CMS 209 signed and dated by the Laboratory Director (LD) had 3 testing personnel listed (TPA, TP B and TP C). Review of CAP PT records found that TP C did not perform PT in 2017-2018 and TP B during 2018. During an interview on 2/26/2019 at 10:30 AM the TP A confirmed that TP C and TP B missed to participate in PT during the years of reference