Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2126052	(X3) Date Survey Completed 03/19/2019
Name of Provider or Supplier Florida Health Care Medical Group	**Street Address, City, State** 107 Jfk Dr #B, Atlantis, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2099	ENDOCRINOLOGY CFR(s): 493.843(b) Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance. This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, the laboratory failed to achieve a score of 80% on one proficiency testing event . Findings include: Review of the American Proficiency Institute (API) records on 03/19/19 indicated that the laboratory got a 0% for thyroxine on the last testing event of 2017. Five of five samples had unacceptable results. During an interview with testing person A at 10:15 a.m on 03/19/19, he confirmed that they had received a 0% score, and that the wrong units had been used when reporting the results