

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2126902	(X3) Date Survey Completed 08/02/2024
Name of Provider or Supplier Dermatology Associates Of Tallahassee Pa	Street Address, City, State 616 St Rd 13 Ste 8, St Johns, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	Dermatology Associates of Tallahassee was found not to be in compliance with 42 CFR Part 493, Requirements for Laboratories as a result of a recertification survey on 8/2/2024.
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview the laboratory failed to perform competency evaluations on 3 (#A, #B, and #C) out of 3 Testing Personnel performing the moderately complex Mycology and Parasitology testing. Findings Include: Review of staff competency evaluations showed no competency evaluations performed on Testing Personnel #A, #B, and #C who performed the moderately complex Mycology and Parasitology testing. During the interview on 8/2/24 at 9:15 AM, laboratory staff confirmed that there were no competencies for Testing Persons #A, #B, and #C.</p>
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to verify the accuracy of KOH (potassium hydroxide) testing at least twice a year in 2023. Findings</p>

Include: Review of peer review records showed no documentation of twice annual verification of KOH testing in 2023. Interview with the laboratory staff on 8/2/24 at 9:14am confirmed that no peer review was performed for KOH in 2023.