

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2127673	(X3) Date Survey Completed 01/21/2025
Name of Provider or Supplier Skin Pathology Center Inc	Street Address, City, State 9271 Sw 68th Street, Miami, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was conducted from January 16, 2025 to January 21, 2025. SKIN PATHOLOGY CENTER INC clinical laboratory was not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5601	<p>HISTOPATHOLOGY CFR(s): 493.1273(a)(f)</p> <p>(a) As specified in 493.1256(e)(3), fluorescent and immunohistochemical stains must be checked for positive and negative reactivity each time of use. For all other differential or special stains, a control slide of known reactivity must be stained with each patient slide or group of patient slides. Reactions of the control slide with each special stain must be documented.</p> <p>This STANDARD is not met as evidenced by: Based on observation, record review, and staff interview the laboratory failed to document quality control activities in the Histopathology laboratory for two random samples in April and August of 2023, out of two years (2023 and 2024) reviewed. The findings included: 1-Review of the daily Skin Pathology Center (SPC) Quality Control (QC) Form for slides dated 04/10/2023 and signed by the laboratory director (Dermatopathologist) on 04/11/2023, did not include positive or negative controls for the random sample received on 04/10/2023 and reported on 04/12/2023. 2- Review of the daily SPC QC Form revealed that there were no QC records, positive or negative controls, for the random sample received on 8/22/2023 and reported on 08/29/2023. 3- Review of the daily "SPC ACCESSION LOG" listed the random patients on 04/3 /2023 and 8/21/2023. 4--Review of the SPC Daily QC procedure stated in step 3. The laboratory will maintain the records of the QC sheets..." 5-Interview on 01/16/2024 at 02:42 PM with the Laboratory Director confirmed that the QC records were missing for two random patients reviewed.</p>