

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2134418	(X3) Date Survey Completed 06/16/2021
Name of Provider or Supplier Quality Pathology Group Inc	Street Address, City, State 260 Nw Peacock Blvd, Suite 101, Port Saint Lucie, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was conducted on June 16, 2021. Quality Pathology Group Inc of Gulfstream Eye clinical laboratory was not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview, the laboratory failed to ensure the Gill 3 Hematoxylin stain was not expired prior to patient testing from 04/02/2020 to 06/16/2021. Findings: Observations made during a tour of the laboratory on 06/16/2021 at 9:20 AM, showed the bottle of Gill 3 Hematoxylin stain expired on 04/01/2020. According to the Clinical Laboratory Improvement Amendments (CLIA) Application for Certification signed and dated by the Laboratory Director on 06/09/2021, the laboratory had an estimated annual test volume of 50 tests per year. On 06/16/2021 at 9:22 AM, the Laboratory Director said the expired Hematoxylin stain was an oversight and would be replaced with a new bottle of stain.</p>