

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 10D2138724	<b>(X3) Date Survey Completed</b> 06/06/2019
<b>Name of Provider or Supplier</b> Oneblood Inc	<b>Street Address, City, State</b> 615 N Bonita Ave, Panama City, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2159</b>	<p><b>ABO GROUP AND D(RHO) TYPING</b> CFR(s): 493.859(d)</p> <p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of CAP ( College of American Pathologists) proficiency testing records and interview with the laboratory manager, the laboratory failed 1 out of 3 events for ABO/RHO test during the year 2018. Findings included: Review of CAP proficiency records revealed that the laboratory failed the ABO/RHO second event of 2018 with a 0 % score due to late submit of results. During an interview on 6 /6/19 at 11:30 AM, the laboratory manager confirmed that the laboratory saved the results on file but failed to submit on time the event of reference.</p>
<b>D2169</b>	<p><b>UNEXPECTED ANTIBODY DETECTION</b> CFR(s): 493.861(c)</p> <p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of CAP ( College of American Pathologists) proficiency testing records and interview with the laboratory manager, the laboratory failed 1 out of 3 events for Unexpected Antibody Detection test during the year 2018. Findings included: Review of CAP proficiency records revealed that the laboratory failed the Unexpected Antibody Detection second event of 2018 with a 0 % score due to late</p>

	<p>submit of results. During an interview on 6/6/19 at 11:30 AM, the laboratory manager confirmed that the laboratory saved the results on file but failed to submit on time the event of reference.</p>
<p><b>D2178</b></p>	<p><b>COMPATIBILITY TESTING</b> CFR(s): 493.863(c)</p> <p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of CAP ( College of American Pathologists) proficiency testing records and interview with the laboratory manager, the laboratory failed 1 out of 3 events for Compatibility Testing during the year 2018. Findings included: Review of CAP proficiency records revealed that the laboratory failed the Compatibility Testing second event of 2018 with a 0 % score due to late submit of results. During an interview on 6/6/19 at 11:30 AM, the laboratory manager confirmed that the laboratory saved the results on file but failed to submit on time the event of reference.</p>
<p><b>D2187</b></p>	<p><b>ANTIBODY IDENTIFICATION</b> CFR(s): 493.865(c)</p> <p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of CAP ( College of American Pathologists) proficiency testing records and interview with the laboratory manager, the laboratory failed 1 out of 3 events for Antibody Identification test during the year 2018. Findings included: Review of CAP proficiency records revealed that the laboratory failed the Antibody Identification test second event of 2018 with a 0 % score due to late submit of results. During an interview on 6/6/19 at 11:30 AM, the laboratory manager confirmed that the laboratory saved the results on file but failed to submit on time the event of reference.</p>