

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2144349	(X3) Date Survey Completed 09/09/2019
Name of Provider or Supplier Bernhardt Laboratories Inc	Street Address, City, State 950 River Edge Court, Longwood, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was conducted on September 9, 2019. Bernhardt Laboratories Inc was found not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5645	<p>CYTOLOGY CFR(s): 493.1274(d)(3)</p> <p>(d) Workload limits. The laboratory must establish and follow written policies and procedures that ensure the following: (d)(3) The laboratory must maintain records of the total number of slides examined by each individual during each 24-hour period and the number of hours spent examining slides in the 24-hour period irrespective of the site or laboratory.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to document and maintain records of the number of hours spent examining cytology slides during each 24 hour period. Findings: Review of the laboratory's workload records showed that the laboratory failed to record the amount of time spent examining cytology (Pap Smear) slides during each 24 hour period from 6/2/19 through 9/9/19. During an interview on 9/9/19 at 11:43 AM, the Laboratory Director acknowledged that she did not record the amount of time spent examining Pap Smear slides.</p>