

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2147363	(X3) Date Survey Completed 01/06/2026
Name of Provider or Supplier Advanced Dermatology And Cosmetic Surgery	Street Address, City, State 329 East Olympia Ave, Punta Gorda, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Advanced Dermatology and Cosmetic Surgery on 1/06/2026. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. The Standard deficiency cited was as follows:
D5801	<p>TEST REPORT CFR(s): 493.1291(a)</p> <p>(a) The laboratory must have an adequate manual or electronic system(s) in place to ensure test results and other patient-specific data are accurately and reliably sent from the point of data entry (whether interfaced or entered manually) to final report destination, in a timely manner. This includes the following: (a)(1) Results reported from calculated data. (a)(2) Results and patient-specific data electronically reported to network or interfaced systems. (a)(3) Manually transcribed or electronically transmitted results and patient-specific information reported directly or upon receipt from outside referral laboratories, satellite or point-of-care testing locations.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to have accurately sent patient Histopathology results from the manually documented results to the final report destination for one (#2883620) of three (#5995851, 2883620, and 5384572) patient records. Findings included: 1. The Histopathology document used to document the patient testing for #2883620, dated 10/24/2025, indicated the testing location as another laboratory. This document was scanned into the patient's permanent final record. 2. Testing Personnel C confirmed on 1/06/2026 at 1:20 p.m., the manually documented results reported into the patient's final report was not accurate as the address was for another laboratory.</p>