

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2152321	(X3) Date Survey Completed 11/06/2024
Name of Provider or Supplier Lakeside Dermatology	Street Address, City, State 727 Us Hwy 27 S, Sebring, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Lakeside Dermatology on 11/06/2024. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiency:
D3011	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on observation, review of policies and procedures and Safety Data, and interview with compliance staff, the laboratory failed to follow safety policies and procedures for chemical and biohazard material for one of one gallon of reagent stored outside of the flammable cabinet. Findings included: On 11/06/2024 at 11:10 am, a beverage cup with a straw and liquid inside the cup was on the lab workbench. Observation of the entrance door to the lab revealed a biohazard, drinking prohibited in this area sign. During the tour, on 11/06/2024 at 11:25 am, a gallon container of 100% Reagent Alcohol was on the floor. The product label revealed the alcohol was flammable and should be stored "locked up." Review of the manufacturer Safety Data Sheet for the 100% Reagent Alcohol revealed, "Highly flammable liquid and vapor" and "Store in an approved Flammable Liquids storage area." Review of the lab Policy and Procedure Manual, approved by the Lab Director in January 2024, under Safety Summary revealed no drinking in the laboratory and ensure no flammable items are stored in the facility except as operations demand. Interview with compliance personnel A on 11/06/2024 at 11:25 am confirmed the beverage should not be present</p>

in the lab. Interview with compliance personnel A and B on 11/06/2024 at 11:33 am confirmed the 100% Reagent Alcohol was not in the flammable cabinet but should be stored in the flammable cabinet.