

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D2152572	<b>(X3) Date Survey Completed</b>  09/20/2022
<b>Name of Provider or Supplier</b>  Delray Dermatology Llc	<b>Street Address, City, State</b>  550 Se 6th Ave Suite 100, Delray Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced recertification survey was conducted on 9/20/22 at Delray Dermatology LLC, a clinical laboratory in Delray Beach, Florida. Delray Dermatology LLC is not in compliance with Code of Federal Regulations (CFR) 42, Part 493, Laboratory Requirements.
<b>D3011</b>	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on observation, record review, and interview the laboratory failed to store 100% Reagent alcohol according to the manufacturer's instructions. The findings included: During a tour of the laboratory on 09/20/22 at 1:00 p.m., 4 boxes that contained 4-1 gallon bottles of 100% Reagent alcohol was seen stacked up beside of the cryostat. Also observed was a 5 gallon and 10 gallon waste container in front of the boxes of 100% Reagent alcohol that contained 100% Reagent alcohol. Review of Mercedes Scientific safety data sheet for 100% Reagent alcohol revealed that the storage conditions was to "Store in an approved Flammable Liquids storage area." Interview on 09/20/22 the Laboratory Manager confirmed that the 100% Reagent alcohol was stored outside of the flammable cabinet.</p>
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p>

This STANDARD is not met as evidenced by:

Based on record review and interview the laboratory failed to verify the accuracy of testing at least twice a year for 1 (2021) out of 2 years (2020-2022) reviewed for Histopathology. The findings included: Review of Quality Assurance (QA) peer reviews from 07/2020 to 09/2022 revealed that there was no documentation of any performed in 2021. Interview on 09/20/22 at 1:00 PM the Laboratory Manager confirmed that there were no peer reviews to verify the accuracy of testing in 2021.