Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 10D2168554	**(X3) Date Survey Completed** 11/26/2019
Name of Provider or Supplier Genesiscare Usa Of Florida Llc	**Street Address, City, State** 20601 E Dixie Hwy, Ste 330, Aventura, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	An initial certification survey conducted on 11-26-19, found the 21st Century Oncology LLC /DBA/ Florida Precision clinical laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D2128	HEMATOLOGY CFR(s): 493.851(e) (1) For any unsatisfactory analyte or test performance or testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure. (2) For any unacceptable analyte or testing event score, remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event. This STANDARD is not met as evidenced by: Based on proficiency testing record review and staff interview, the laboratory failed to receive a passing score of 80 percent for the White Blood Cell Differential analyte for the second event of American Proficiency Institute (API) in 2019 . Findings include : A review of the API proficiency testing record indicated the laboratory received a score of 16 percent for the White Blood Cell Differential analyte for the second event in June 30th , 2019 . On 11-26-19 at 12:00 pm , the testing person A and office consultant confirmed that laboratory had received a score of 16 percent for the White Blood Cell Differential analyte for the 2nd event in 2019 .