

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 10D2170233	<b>(X3) Date Survey Completed</b> 04/01/2025
<b>Name of Provider or Supplier</b> Endocrinology Of Central Florida	<b>Street Address, City, State</b> 141 Victoria Commons Blvds, Deland, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A recertification survey was conducted on April 1, 2025. Endocrinology of Central Florida was found not in compliance with 42 CFR 493, requirements for clinical laboratories.
<b>D5473</b>	<p><b>CONTROL PROCEDURES</b> CFR(s): 493.1256(e)(2)(g)</p> <p>(e)(2) Each day of use (unless otherwise specified in this subpart), test staining materials for intended reactivity to ensure predictable staining characteristics. Control materials for both positive and negative reactivity must be included, as appropriate.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the facility failed to maintain daily quality control (QC) slide documentation for Cytology stains for two days out of three months reviewed in 2024 (8/7/24 &amp; 9/4/24). Findings include: Review of QC documentation showed no record for the quality of the Cytology stains used during Fine Needle Aspiration Biopsy on 8/7/24 (3 patients tested) and 9/4/24 (3 patients tested). Interview with the Nurse Manager on 4/1/25 at 10:35 AM confirmed that the documentation was missing.</p>