

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2174100	(X3) Date Survey Completed 02/22/2022
Name of Provider or Supplier Matthew Taub Md Pa	Street Address, City, State 4700 Sheridan Street Suite M, Hollywood, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey conducted on 02/22/2022 found that MATTHEW TAUB MD PA clinical laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D2122	<p>HEMATOLOGY CFR(s): 493.851(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of proficiency testing (PT) records and staff interview, the laboratory received an unsatisfactory score in second event of 2021 for the specialty of Hematology. Findings include: Review of American Proficiency Institute (API) PT records revealed a score of 76 % for White Blood Cell Differential, 20% for Red Blood Cells, 20 % Hematocrit, 0% Hemoglobin, 0 % White Blood Cells, 20 % Platelets tests resulting in an overall score of 22 % in the second event of 2021. During an interview on 02/22/2022 at 11:30 AM, Technical Consultant B confirmed the proficiency testing failure.</p>