

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2184905	(X3) Date Survey Completed 01/12/2023
Name of Provider or Supplier Dermatology Group Of Florida Pa	Street Address, City, State 9271 Sw 68th Street Division 2, Miami, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey conducted on 01/12/2023 found the DERMATOLOGY GROUP OF FLORIDA PA clinical laboratory not in compliance with 42 CFR Part 493, Requirements for Laboratories
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on lack of records and laboratory consultant interview, the laboratory failed to ensure the twice a year accuracy verification for histopathology testing for two out of two years reviewed (2021 and 2022). Findings include: -The laboratory failed to have documentation that have performed peer review during 2021 and 2022. -During an interview on 01/12/2023 at 10:30 AM, the laboratory consultant confirmed that the laboratory failed to do twice a year accuracy verification for histopathology during 2021 and 2022.</p>