

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2187778	(X3) Date Survey Completed 02/29/2024
Name of Provider or Supplier Cancer Specialists, Llc	Street Address, City, State 80 Pinnacle Dr Suite 700, Palm Coast, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was conducted on 2/29/24. Cancer Specialists, Llc was found to be not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5415	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(c)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following: (1) Identity and when significant, titer, strength or concentration. (2) Storage requirements. (3) Preparation and expiration dates. (4) Other pertinent information required for proper use.</p> <p>This STANDARD is not met as evidenced by: Based on observation and staff interview, the facility failed to ensure the glass jars used for the Quick III Hematology Stains for staining peripheral blood smears were labeled with the identity of the stain, preparation, and expiration date. The findings include: During the tour of the laboratory on 2/29/24 at 11:15am, three jars with no identifying labels were observed on the counter in the hematology testing area with light blue, orange, and purple colored liquids inside. The interview with the testing person on 2/29/24 at 11:20am confirmed the jars were used for making blood smears using the Quick III procedure for performance of cell differentials. .</p>