

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2215918	(X3) Date Survey Completed 11/24/2025
Name of Provider or Supplier Academic Alliance In Dermatology Inc	Street Address, City, State 6600 University Parkway Ste 302, Sarasota, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3011	FACILITIES CFR(s): 493.1101(d) Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials. This STANDARD is not met as evidenced by: Based on observation, review of purchase order dated 9/03/2025, and interview, the laboratory failed to ensure safety procedures were observed to ensure protection from physical and chemical hazards. This is a new citation. Findings included: 1. On 11/24/2025 at 9:37 a.m., during a tour of the laboratory the surveyor observed a jug of 100% Dehydrant with "Flammable Hazard " labeling seated on the laboratory counter and not in the required safety cabinet. 2. On 11/24/2025 at 9:37 a.m. during a tour of the facility, no fume hood was observed in the laboratory for the chemicals used in Histopathology staining and processing. A Purchase Order dated 9/03/2025 was presented showing a Laconco Fume Absorber had been ordered on 9/03/2025, however had not been delivered or installed. 3. The Regional Director of Operations confirmed on 11/24/2025 at 9:52 a.m. via phone call the 100% Dehydrant was to be stored in the safety cabinet to protect from possible fire and that a fume hood to protect from chemical hazards was required. 4. The Laboratory Director of record was not available during the 11/24/2025 visit.