

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D2256885	<b>(X3) Date Survey Completed</b>  08/24/2022
<b>Name of Provider or Supplier</b>  Pediatric Associates	<b>Street Address, City, State</b>  16250 Nw 57th Ave, Miami Lakes, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An initial certification survey conducted at PEDIATRIC ASSOCIATES on 08/24 /2022 found the clinical laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories.
<b>D2122</b>	<p>HEMATOLOGY CFR(s): 493.851(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of proficiency testing (PT) records and staff interview, the laboratory failed to receive a passing score for the second event of 2022 for the specialty of Hematology. Findings include: Review of American Proficiency Institute (API) PT Performance Summary revealed unsatisfactory scores for the Center for Medicare &amp; Medicaid Services (CMS) reportable analytes of White Blood Cell Differential 20%, Erythrocyte Count 20%, Hematocrit 0%, Hemoglobin 0%, Leucocyte Count 20% and Platelet Count 0% resulting in an overall score of 7% for the specialty of Hematology in the second event of 2022. During an interview on 08/24/2022 at 01:30 PM, the Area Supervisor confirmed the proficiency testing failure.</p>