

| | | |
|--|--|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 10D2256885 | (X3) Date Survey Completed 01/12/2026 |
| Name of Provider or Supplier Pediatric Associates | Street Address, City, State 16250 Nw 57th Ave, Miami Lakes, FL | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|---|
| D0000 | An announced CLIA recertification survey was conducted at PEDIATRIC ASSOCIATES from December 18, 2025 to January 12, 2026. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Standard deficiencies cited are as follows: |
| D2007 | <p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1)</p> <p>(b)(1) The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the Laboratory failed to have all testing personnel (TP) rotate through the testing of Proficiency Testing (PT) in the Specialty of Chemistry for five out of five events reviewed 2024 (2nd, 3rd) and 2025 (1st, 2nd, 3rd). Findings included: 1- Review of form CMS-209 signed and dated by the Laboratory Director (LD) on 01/08/2026 revealed that the laboratory had nine Testing Personnel (TP1, TP2, TP3, TP4, TP5, TP6, TP7, TP8, and TP9) listed. 2- Review of personnel files revealed that TP1, TP2, TP3, TP4, TP6 and TP7 were TP in 2024 and 2025. TP5, TP8 and TP9 were TP in 2024. 3- Review of the policy LAB 410 Proficiency testing effective date 05/05/2024 stated "Analytical Phase - All Laboratory testing personnel will rotate testing events equally, using the API test schedule table (including part time personnel)." 4- Review of American Proficiency Institute (API) PT records for Chemistry (2nd and 3rd) 2024 and (1st, 2nd, and 3rd) 2025 revealed that TP1 performed all five events. 5- During interview on 01/08/2026 at approximately 12:10 PM, the Laboratory Lead confirmed that the laboratory failed to rotate Testing Personnel for Chemistry in 2024 and 2025.</p> |