

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2278930	(X3) Date Survey Completed 10/27/2023
Name of Provider or Supplier Sawgrass Pediatric Partners Llc	Street Address, City, State 7895 N University Drive, Parkland, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An initial certification survey conducted at SAWGRASS PEDIATRIC PARTNERS LLC from 10/23/2023 to 10/27/2023 found the clinical laboratory not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D5411	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(a)</p> <p>Test systems must be selected by the laboratory. The testing must be performed following the manufacturer's instructions and in a manner that provides test results within the laboratory's stated performance specifications for each test system as determined under 493.1253.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to read throat cultures (TC) at 24 hours per their policy for 14 out of 155 patients reported since 05/19/2023. Findings included: Review of the policy "DOCUMENTATION OF THROAT CULTURES" revealed that: "MD READS 24 HR AND IS ONLY DOCUMENTED IN LOG. MD TO READ 48 HR THROAT CULTURE, DOCUMENT IN LOG BOOK, ALSO ENTER RESULTS IN EHR." Review of "DAILY RAPID STREP/TC LOG" revealed the days and patients that results were not read at 24 hours, instead they were read at 48 hours: 05/19/2023: one patient 05/26/2023: six patients 07/23 /2023: seven patients During an interview on 10/23/2022 at 12:30 PM, the laboratory director confirmed that the laboratory failed to read 14 TC at 24 hours as per their policy.</p>