Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 10D2288565	**(X3) Date Survey Completed** 10/06/2025
Name of Provider or Supplier Sunshine State Dermatology And Skin Cancer	**Street Address, City, State** 12497 S Tamiami Trail, Unit 1, North Port, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	An announced CLIA recertification survey was conducted at Sunshine State Dermatology and Skin Cancer Center, Inc on 10/06/2025. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Standard deficiencies cited are as follows:
D5219	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(2) (c)(2) Any test or procedure listed in subpart I of this part for which compatible proficiency testing samples are not offered by a CMS-approved proficiency testing program. This STANDARD is not met as evidenced by: Based on record review and interview, two of two Testing Personnel (TP-A and TP-B) failed to perform twice annual Parasitology (Scabies) verification of accuracy testing from 3/2024-10/2025. Findings include: 1. Review of the CMS-209 Laboratory Personnel Report signed and dated by the Laboratory Director 9/30/2025, listed two Testing Personnel (TP-A and TP-B) who performed Parasitology (Scabies). Both TP-A and TP-B had been listed as performing Parasitology (Scabies) on the CMS-209 Laboratory Personnel Report signed and dated by the Laboratory Director 3 /07/2024. 2. Scabies proficiency testing worksheet recorded only one proficiency record for TP-B dated 4/28/2025. There failed to be any Scabies proficiency testing worksheet for TP-A. 3. The Compliance Manager on 10/06/2025 at 12:05 p.m. confirmed the laboratory failed to perform twice annual Parasitology (Scabies) verification of accuracy testing from 3/20204-10/2025.