

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2299323	(X3) Date Survey Completed 07/09/2026
Name of Provider or Supplier Florida Center For Dermatology Laboratory Ii	Street Address, City, State 3 Shircliff Way Suite 630, Jacksonville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Florida Center for Dermatology Laboratory II on 7/9/2026. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D3007	FACILITIES CFR(s): 493.1101(b) (b) The laboratory must have appropriate and sufficient equipment, instruments, reagents, materials, and supplies for the type and volume of testing it performs. This STANDARD is not met as evidenced by: Based on observation and interview, the laboratory failed to ensure that it maintained appropriate and sufficient equipment necessary for accurate and reliable test system operation by failing to have a thermometer available in the Mohs laboratory to monitor required environmental conditions. Findings include: Observation on 07/09 /2026 at 9:15am revealed that no thermometer was present in the Mohs laboratory. Although the laboratory's policy and the "Room Temp/Humidity Log" required the daily recording of room temperature and humidity, the equipment necessary to obtain these measurements was not available on the premises. In an interview on 07/09/2026 at 10:00 AM, Staff #A (Mohs Technician) confirmed that there was no thermometer currently in the laboratory. Staff #A stated that the thermometer is "usually here" and further noted that "it gets warm in the laboratory usually around 2 PM."