Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0257657	(X3) Date Survey Completed 01/24/2018
Name of Provider or Supplier Northside Gwinnett Hospital	**Street Address, City, State** 1000 Medical Center Blvd, Lawrenceville, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A complaint survey was performed on February 23, 2017. Gwinnett Medical Center was not in compliance with the 42 CFR Part 493, Requirements for Laboratories. The Conditions at 42 CFR 493.801- Enrollment and testing of samples. Laboratory A is Gwinnett Medical Center (11D0257657) 1000 Medical Center Boulevard Lawrenceville, Georgia 30046. Laboratory B Gwinnett Medical Center-Duluth (11D0256605) 3620 Howell Ferry Road Duluth, Georgia 30096. Laboratory B sent CAP ID3-A 2016 Influenza A, Influenza B, and RSV Proficiency Testing (PT) samples to Laboratory A for analysis since Laboratory A used all of their PT samples. Laboratory A engaged in inter-laboratory communications with Laboratory B pertaining to the results of proficiency testing samples prior to the date by which Laboratory A was required to report proficiency testing results to the College of American Pathologist (CAP). Laboratory A reported the results of Laboratory B's PT samples to CAP in place of Laboratory A's own incorrect results.