

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  11D0258633	<b>(X3) Date Survey Completed</b>  06/06/2019
<b>Name of Provider or Supplier</b>  Dermatology Consultants Pc	<b>Street Address, City, State</b>  3280 Howell Mill Road, Nw, Suite 101, Atlanta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on June 6, 2019. The laboratory was found to be not in compliance with all applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
<b>D3011</b>	<p><b>FACILITIES</b> CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation and an interview with office manager, the laboratory failed to ensure safety procedures for proper storage of medicine, patient samples, and laboratory supplies in a designated area. The findings include: 1. An observation of one of the laboratory refrigerators revealed that beverages and food items were being stored in the same area as the Cosentyx prefilled injectable, previously tested patient samples, and other laboratory supplies, which affects the integrity, quality, and purity of the products. 2. The office manager confirmed on June 6, 2019 at 10:45 AM in the laboratory that beverages and food were being stored in the same refrigerator as laboratory supplies and other medicines.</p>