Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 11D0259153	**(X3) Date Survey Completed** 07/09/2019
Name of Provider or Supplier Atlanta Center For Medicine	**Street Address, City, State** 2801 North Decatur Rd Ste 300, Decatur, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	An initial Clinical Laboratory Improvement Amendments (CLIA) survey was completed on July 9, 2019. The laboratory was not in compliance with all applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D6029	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(11) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(11) Ensure that prior to testing patients' specimens, all personnel have the appropriate education and experience, receive the appropriate training for the type and complexity of the services offered, and have demonstrated that they can perform all testing operations reliably to provide and report accurate results. This STANDARD is not met as evidenced by: Based on testing personnel (TP) document review and staff interview, the laboratory director (LD) failed to ensure TP had appropriate training for the type and complexity of the services offered as required. Findings include: 1. TP document review revealed Staff #1 (CMS 209) had no documentation of training for the Sysmex XS 1000i hematology analyzer in 2019. 2. An interview with the laboratory supervisor in a medical office at approximately 5:00 p.m. confirmed there was no Sysmex training documentation for Staff #1 (CMS 209) at the time of survey.